

DEPARTMENT OF FORENSIC MEDICINE & TOXICOLOGY  
A.I.I.M.S. NEW DELHI

R.E. No. 01/FMT/17-18

Dated: 26.04.2017

Subject: **Short Rate Enquiry for purchase of Autopsy Electric Saw, Qty. 01 (Specification attached).**

**Terms & conditions**

1. The quotation should be addressed to the Prof. & Head, Dept. Forensic Medicine & Toxicology, Room No. 302, 2<sup>nd</sup> Floor, Near New Mortuary Complex, New Delhi and reach this office on or before **08.05.2017** duly sealed marked R.E. No. 01/FMT/17-18 dated 26.04.2017. Catalogue/broachers / description with terms & condition must be attached with the technical bid.
2. Rate Enquiry should be submitted in the two bid system containing two parts as detailed below. EMD/Bid Security Money of Rs. 5000/- (Demand Draft) in favour of "Director AIIMS" must be attached with techno-commercial bid. (Technical bid without EMD will be accepted and treated as disqualified).  
PART-I : TECHNO-COMMERCIAL BID IN ONE SEALED COVER  
PART-II : PRICE BID/FINANCIAL BID IN ONE SEALED COVER
3. Both the sealed envelopes should then be put in one outer cover indicating thereon:
  - i) Reference R.E. No.....
  - ii) R.E. Regarding.....
  - iii) Due date of Submission Rate Enquiry.....

PLEASE NOTE THAT PRICES SHOULD NOT BE INDICATED IN THE TECHNO-COMMERCIAL BID.

NOTE: RATE ENQUIRY SUBMITTED WITHOUT FOLLOWING TWO BID SYSTEM PROCEDURE AS MENTIONED ABOVE WILL BE SUMMRILY REJECTED.

To : \_\_\_\_\_  
\_\_\_\_\_

Dr. Millo Tabin

Prof. & Faculty Incharge Stores

**Dr. MILLO TABIN**

Professor

Faculty Incharge - Store  
Dept. of Forensic Medicine  
AIIMS, New Delhi-110029

## Technical Specifications of Autopsy Electric Saw

### Autopsy Electric Saw:

Model should be specifically designed for Autopsy Procedure.  
Handle near blade provides good control and balance.  
It should be splash proof and have a safety reset system.  
Protection fuse.

Oscillating speed range:	12000-25000 osc/min
Motor protection:	Overload cut-out, auto reset
Mode of operation:	Continuous
Cable length:	3-4 metres
Length:	270-350 mm
Weight:	less than 2 kg
Sound level:	less than 80 DB

Supply voltages:	220-240 V
Output to saw:	40-50 V
Supply frequency:	50-60 Hz
Weight:	less than 3.5 Kg
Certification:	CE certified

It should be supplied with 5 blades compatible for Autopsy saw and other accessories.

The staff will have to be trained for the use of Autopsy Saw.