

**STORE SECTION (CNC)  
(AIIMS)**

**T. No.211/CNC/Cardio/2023-2024/St.**

**Dated: 08.05.2024**

**Subject: Amendment in the Buy-back equipment, Reserve price & Change of Turnkey location & area-reg.**

\*\*\*\*\*


With reference to Additional representation letter submitted by M/s.Wipro GE Healthcare Pvt. Ltd. vide letter No.WGE/AIIMS/2024/SB/002 dated 24.03.2024 i.e.

- 1.The X-ray tube should have anode heat storage capacity of at least 3.5 MHU or more.
2. Payment Terms should be 100% through Inland Letter of Credit. At least main the GeM Payment terms which is 80% on delivery and 20% on CARC.
3. During the warranty & CAMC period, we will maintain 95% Uptime warranty for the equipment. However, if downtime exceeds 5% then the warranty & CAMC period shall be proportionally extended by double of the period for such equipment. On receipt of the notice from the consignee, we will replace the defective parts within least possible time. Remote facility should be provided by the Customer. So, we request you to kindly remove the penalty of 0.25% of the total cost of the equipment per day will be liable for the excess downtime period but not more than 5% of the equipment cost. Due to being high end equipment, we can't provide the replacement unit. We assure you; we will maintain 95% Uptime warranty during the warranty & CAMC Period for the equipment. However, if downtime exceeds 5% then the warranty & CAMC period shall be proportionally extended by double of the period for such equipment. So, please remove this clause.

As same request has been submitted by the same firm during pre-bid meeting and after discussion with TSEC, Finance Division, as commercial terms mentioned in the bid which is as per AIIMS PURCHASE MANUAL, the request for point No.2&3 have had not been considered by the TSEC. The same amendment request again sent by the firm which has not been considered by the user department including point No.01 (As per Note No.49 of Eoffice No.3241523). as point No.1 has already been diluted to 3.0 MHU or more in order to comply and participate more firms). The following amendment is being carried out as mentioned below:

<b>Existing specifications/description</b>	<b>Amended specifications/description</b>
<b>Turnkey location:</b> Cath Lab-III (The approximate area is 1400 sqft.) Specifications/content will remain same as mentioned in the revised specifications (after pre-bid meeting)	<b>Turnkey location:</b> Cath Lab-II turnkey area is 586 sq feet but total turnkey area is 1400sqft. including patient care corridor, doctor movement area, Seminar area, nurse and paramedical staff area. Specifications/content for turnkey will remain same as mentioned in the revised specifications (after pre-bid meeting). The vendors must visit site before quote the rates/uploading the bid.
<b>Buy-back:</b> The bidder must visit site (Cath Lab.-III, Ground Floor; CNC, AIIMS, New Delhi-29), see existing Single plane Cardiovascular catheterization with digital substration Angiography equipment (Make: GE Model: Innova 2100) installed in the year 2010. The minimum reserve price (As per depreciation policy of companies act-2013) for buy-back of this equipment is Rs.16,42,414/-. The vendors are advised to see the old equipment with prior permission of HOD, Cardiac Radiology/Cardiology and quote the buy-back price (in break-up) positively. The buy-back price (minus) will be considered for ranking purpose.	<b>Buy-back:</b> The bidder must visit site (Cath Lab.-II, Ground Floor, CNC, AIIMS, New Delhi-29), see existing FD-10 (Sr.No.287) equipment (Make: Philips) installed in the 28.12.2011. The minimum reserve price (As per depreciation policy of companies act-2013) for buy-back of this equipment is Rs.18,98,068/-. The vendors are advised to see the old equipment with prior permission of HOD, Cardiac Radiology/Cardiology and quote the buy-back price (in break-up) mandatorily. The buy-back price (minus) will be considered for ranking purpose.

**The other specifications/content/terms & conditions of the bid uploaded (after pre-bid meeting) will remain same. The bidder must see and quoted their bid accordingly.**

  
**STORE OFFICER (CNC)**