

**ALL INDIA INSTITUTE OF MEDICAL SCIENCES
ANSARI NAGAR, NEW DELHI-29
STORE SECTION (CNC)**

Dated: 10/03/2025

TENDER CORRIGENDUM

The Tender No.129/CNC/NP/24-25/ST. with GeM Portal bid vide No.GEM/2025/B/5899342 dated 07/02/2025 was published in GeM Portal (e-Procurement) for purchase of **Automated Slide Printer-01No.** for the department of NP, CNC, AIIMS, NEW DELHI. Meanwhile, a firm DK Lab Solutions has submitted a representation dated 14/02/2025. Upon the request of the potential vendors & recommendations of TSEC, few amendments have been carried out in the published specifications. Further, it is also informed that bid end date was 07/03/2025, which has been already expired; accordingly, extension of the bid end date & bid opening date and amendment in specifications is being carried out as under:

Description	Existing	Amended
Bid submission end Date and Time	07-03-2025	20-03-2025
Bid Opening Date and Time	07-03-2025	20-03-2025

- The following amendments have been done by the TSEC.

S. No.	Offered Specifications	Changed Specifications	Remarks
7.	Laser slide printing resolution shall be at least 600 dpi and not more than 600 dpi and printing should be monochromatic black.	Laser slide printing resolution shall be at least 600 dpi and printing should be monochromatic black.	It better to have higher dpi for better print quality. Company specific
12.	The barcode scanner should be mounted under the left front edge of the screen of the slide printer & it should use a LED light source requiring no special eye protection during use of the scanner.	The instrument should an inbuilt barcode scanner or barcode scanner at convenient position & it should use a LED light source requiring no special eye protection during use of the scanner.	Kindly remove Company specific point
16.	Should have two collection system should be integrated with laser slide printer to collect 10 slides in each collection system after slides are printed with collection capacity	Should have collection system to be integrated with laser slide printer to collect at least 20 slides in collection system after slides are printed with collection capacity up to 20 slides	Larger collection capacity is always recommended for unattended operations Company specific point
23.	Each template shall contain not less than 9 data fields.	Each template shall contain not less than 8 data fields.	Normally 6 data fields are sufficient for any format For better participation

All the prospective bidders are advised to see above amendment and submit their bid accordingly. All other published specifications' Terms and Conditions will remain same.

This is for information to all concerned.

[Signature]
Store Officer (CNC)