

STORE SECTION (D.O.)

ALL INDIA INSTITUTE OF MEDICAL SCIENCES
ANSARI NAGAR, NEW DELHI - 110029


Ref. No. XX-224/SO(DO)/Anesth./2025-26/M&E

Dated: 17.03.2026

It is informed that the tender bearing no. XX-224/SO (DO)/Anesth./2025-26/M&E and CPP tender I.D. No. 2026_AIMSD_897710_1, procurement of Advanced Dialysis Machine (With SLED facility), was published through CPP portal, has some amendment in technical specification & dates of tender has been extended, as per TSEC recommendation and detail given below:

Sl. No.	DETAIL	EXISTING	AMENDED/EXTENDED
1.	Bid submission END Date	17.03.2026	27.03.2026
2.	Bid Opening Date	18.03.2026	28.03.2026
Sl. No	Existing Specification	Amended read as	
1	a) Flow rate range: 50-600 ml/min with 5 ml/min increments	Flow rate range: 50-600 ml/min with increments up to 10 ml/min	
2	c) Heparinization stop time (before end of treatment) between 0-23 hrs and should be user adjustable	Heparin stop time should be user adjustable with minimum stop time of a least 10 hours	
3	a) Venous pressure monitoring: Range: -100 to + 400 mmHg	Venous pressure monitoring: Range: - 50 to + 350 mmHg	
9	c) Treatment Time: adjustable up to 22 hrs in 5 min increment	Treatment Time: adjustable up to 10 hours	
14	Prolonged intermittent renal replacement therapy: Should be able to provide SLED/PIRRT for up to 24 hours continuously without need of rinsing	Prolonged intermittent renal replacement therapy: Should be minimum SLED/PIRRT for 12 hours	

All other terms & condition of the tender are same.


17/03/2026
STORE OFFICER (D.O.)
17/3