

**STORE SECTION (D.O.)**

ALL INDIA INSTITUTE OF MEDICAL SCIENCES  
ANSARI NAGAR, NEW DELHI – 110029


**Dated: 03.01.2026**


**Ref. No. XX-142/SO(DO)/Micro./2025-26/M&E**

It is informed that the tender bearing no. XX-142/SO (DO)/Micro./2025-26/M&E and CPP tender I.D. No. 2025\_AIMSD\_884534\_1, procurement of FlowCytometer, was published through CPP portal, has some amendment in technical specification, as per TSEC recommendation and detail given below:

Sl. No.	Existing Specification	Amended read as
1.	<b>Sr. No.1</b> System should be a bench-top flow-cytometer with five simultaneous usable lasers (blue, red, violet, UV and Yellow-Green)	System should be a bench-top flow-cytometer with five simultaneous usable lasers (blue, red, violet, UV and Yellow-Green) with all laser powers>50mW
2.	<b>Sr. No.9</b> System should have the capability for compensation in real-time and also post-acquisition.	System should have the capability for compensation/ or spectral unmixing in real-time and also post-acquisition.
3.		<b>Add:</b> Only firm with valid CDSCO compliance shall be eligible to participate in the bid. Submission of valid CDSCO compliance certification is a mandatory preconditions for participation in the bidding process. <b>(copy of same must be enclosed with Technical Bid)</b>

All other terms & condition of the tender are same.

  
STORE OFFICER (D.O.)

  
31/1/26