



# Thermo Fisher Scientific "Sensititre™ Complete Automated AST System"

- a. System should be able to perform antibiotic sensitivity (TRUE MIC) of all microbial species- Gram negative, Gram positive, Anaerobic, Fungal/Moulds and Mycobacteria
- b. System should be able to take & store digital image of MIC plates for readings
- c. System should have multiple light adjustment mode for brightness and contrast
- d. System should be supplied with preloaded software computer system for interpreting TRUE MIC plates
- e. Software should be able to interpret the antibiotic sensitivity of the test organism for a panel of antibiotics to give Sensitive or Resistant or Intermediate or Invalid
- f. Software should give expert notes when results are not falling in the range
- g. The results should be based on current CLSI guidelines
- h. A nephelometer should be provided for adjusting density of culture suspension as per Mc Farland standard.
- i. Should be ETL safety tested, CE marked, EMC tested, FDA-cleared
- j. Power Requirements - 220/240VAC; 60Hz; 8w
- k. Should be provided with epidemiological module (optional)
- l. Operational Manual and Service manual in English should be included.
- m. Compatible desktop/Laptop and UPS 1KV should be provided along with instrument.

  
**डॉ. गगनदीप सिंह**  
**Dr. GAGANDEEP SINGH, MD, MNAMS**  
 Associate Professor  
 Department of Microbiology  
 All India Institute of Medical Sciences  
 Ansari Nagar, New Delhi-110029

  
**Dr. PRASENJIT DAS**  
**MD., DNB, MNAMS, WHO**  
 Additional Professor  
 Department of Pathology  
 Institute of Medical Sciences  
 All India Institute of Medical Sciences  
 Ansari Nagar, New Delhi-110029

  
**डॉ. निशांत वर्मा**  
**Dr. NISHANT VERMA, MD**  
 Associate Professor  
 Department of Microbiology  
 All India Institute of Medical Sciences  
 Ansari Nagar, New Delhi-110029


  
**डॉ. विलादेव सिद्धार्थ Vilaydeep Sidharth**  
 Associate Professor  
 Department of Hospital Administration  
 All India Institute of Medical Sciences  
 Ansari Nagar, New Delhi-110029


**RChaudhry**


  


**ALL INDIA INSTITUTE OF MEDICAL SCIENCES  
ANSARI NAGAR, NEW DELHI-110029**

1	Item/Type Model No. required along with specification	Sens titre Vizion System
2	Is the item a spar parts attachment or accessory for an existing equipment	New Equipment
3	Name of manufacturers/supplier of the item proposed by the indenter	M/s ThermoFisher Scientific
4	Are they sole manufacturers/distributors of the items	Yes
5	Is there any other item with similar/equivalent specification available in the market to meet the job requirement envisaged. If yes, why the same can't be procured. Demanding officer should bring out comparative functional advantages/cost effectiveness of the recommended item from these offered by other	NO
6	What were the efforts made to locate alternative source of supply or use other substitutes	The items mentioned in col.1 are proprietary of M/s ThermoFisher Scientific and are exclusively distributed by them only. Thus no other supplier provides the said items. Also, for the better reproducibility of results other substitutes could not be used.
7	Why open/limited tender can't be resorted to for locating alternative source.	As the said items are proprietary and no other supplier provides the same therefore open/limited tender was not resorted.
8	Are the proprietary items certifying that the rates are reasonable or not.	Rates are reasonable (certificate attached)
9	Any other justification for procuring item from single source.	Proprietary

  
Signature of Indenter  
(Demanding Officer)



  
Counter Signature  
(Head of the Department)

**Dr. BIMAL KUMAR DAS, MD**  
Professor  
Department of Microbiology  
All India Institute of Medical Sciences  
Ansar Nagar, New Delhi-110029

**Dr. RAMA CHAUDHRY**  
Professor & Head  
Department of Microbiology  
All India Institute of Medical Sciences  
Ansar Nagar, New Delhi-110029

I certify that the items at Sr.No.1 above is required to be procured on the single tender basis as the source of supply is definitely known/the specific brand proposed was the only one meeting our functional requirements and limited tender could be resorted to as the would serve no useful purpose in this particular case.

(Strike out whichever is not applicable)

DECLARATION LETTER

Date: 18<sup>th</sup> Dec 2020

To,  
The Director  
All India Institute of Medical Science Ansari Nagar,  
New Delhi-110029

Kind Attention: Dr Gagandeep Singh

Sub./Tender Ref.: P O Declaration Letter

Dear Sir,

We, Thermo Fisher Scientific India Pvt. Ltd. having registered office at 403-404, Delphi "B" Wing, Hiranandani Business Park, Powai, Mumbai – 400076 are affiliate of Thermo Fisher Scientific Inc. USA and its subsidiaries worldwide who are proven and reputed manufacturers (Thermo Fisher Scientific Microbiology Products) having factories at USA and UK.

We have supplied the Thermo Scientific Sensititre Susceptibility and Identification System on reagent rental basis and not executed any exclusive order for the equipment PAN India, so purchase order can not be provided for the same.

We look forward for your kind consideration and support

For Thermo Fisher Scientific India Pvt. Ltd.

*(Name)*



Dr. Pramod Kabra  
Director- Healthcare Solution India

[www.thermofisher.com](http://www.thermofisher.com)

Registered Office:

Thermo Fisher Scientific India Pvt. Ltd. 403, 404, Delphi 'B' Wing, Powai, Mumbai - 400 076,  
Hiranandani Business Park, India

+91-22-6716 2200 tel  
+91-22-6716 2244 fax  
1800222230 toll free number  
CIN: U73100MH2000PTC126872



Quotation

Ref No: GS/AG/TF/131220/1

Date: 13-12-2020

To,  
Dr Gagandeep  
Department of Mycology  
All India Institute of Medical Science  
New Delhi

Sub: Offer for Sensititre Vizion Imaging System  
Ref No: Through mail

Dear Sir,  
We are pleased to offer following against above said reference.

S.N.	Part.No.	Description	List Price (Rs)	Qty	Total Price (Rs)
1	V2020	Sensititre Vizion Imaging System Make- Thermofisher Scientific	696,912	1	696,912
		Combining automation with visual results, the Vizion System consolidates all offline testing on a single instrument, with the benefits of Expert System software and LIS connectivity. <ul style="list-style-type: none"><li>• Customizable lighting options facilitate reading of all organisms</li><li>• Easy-to-read digital images eliminate manual reading errors</li><li>• Stored images expedite teaching, sharing and decisionmaking</li></ul>			
Ex work Price				INR	696,912
Freight & Insurance				INR	NIL
5% against GST certificate and DSIR letter				INR	34,845.6
Total FOR Destination Price – Seven Lakh Thirty One Thousand Seven Hundred Fifty Eight				INR	731,758

### CMC Charges after completion of 5 years of warranty will be

CMC for 1 <sup>st</sup> Year after 5-year warranty	48,000	1	48,000
CMC for 2 <sup>nd</sup> Year after 5-year warranty	52,000	1	52,000
CMC for 3 <sup>rd</sup> Year after 5-year warranty	56,000	1	56,000
CMC for 4 <sup>th</sup> Year after 5-year warranty	60,000	1	60,000
CMC for 5 <sup>th</sup> Year after 5-year warranty	64,000	1	64,000



GINGER SCIENCE PVT. LTD.

LGF-6-8, Jinkushal Suri Complex, 230-A, Masjid Moth, South Ex-II, New Delhi – 110049

Tel No : 011-26263236 /26253236 | sales@gingerscience.co.in | www.gingerscience.co.in

CIN: U74999DL2014PTC269150


## :Terms & Conditions:

1	Prices	F.O.R
2	Payment Term	Within 30 days
3	Delivery Period	4 to 6 weeks from the date of confirmed purchase order
4	Warranty/Shelf Life	60 months from the date of installation
5	Installation	FOC
6	Training	FOC
7	Validity	30 Days
8	Order to Be Placed:	<b>Ginger Science Pvt. Ltd.</b> 230-A, Masjid Moth, South Ex-II New Delhi – 110049 Tel No: 011-26263236 /26253236 Cell No.: +91- 9599439598 ayush.gupta@gingerscience.co.in sales@gingerscience.co.in <a href="http://www.gingerscience.co.in">www.gingerscience.co.in</a>
9	Additional Document	GST certificate and DSIR letter

Thank you  
Yours Sincerely  
Kind Regards



Ayush Gupta  
Cell No.: +91- 9599439598  
E mail id: ayush.gupta@gingerscience.co.in

  
GINGER SCIENCE PVT. LTD.  
LGF-6-8, Jinkushal Suri Complex, 230-A, Masjid Moth, South Ex-II, New Delhi – 110049  
Tel No : 011-26263236 /26253236 | sales@gingerscience.co.in | [www.gingerscience.co.in](http://www.gingerscience.co.in)  
CIN: U74999DL2014PTC269150

June 27, 2018

**Subject: Sole Source Thermo Scientific™ Sensititre™ System**


To Whom It May Concern:

Please accept this letter as certification that Thermo Fisher Scientific is the sole source manufacturer of the Sensititre System and their quality is certified by CE.

Description
Sensititre MIC Plates w/CE Mark
Sensititre Nephelometer
Sensititre AIM
Sensititre Vizion
Sensititre OptiRead

Sensititre instruments are manufactured for Thermo Fisher Scientific under the Thermo Scientific brand according to specifications and supplied exclusively from Thermo Fisher Scientific.

Sincerely,



Sally Maysent  
Senior Product Manager – AST  
Microbiology  
Thermo Fisher Scientific

