



Proprietary Certificate


We hereby certify to the best of our knowledge and belief that the "CAPSULaser cataract system" is the proprietary product of EXCEL-LENS, Inc., USA with the following unique and proprietary features.

- CAPSULaser laser system uses a "Quantum dot, optically pumped thin disc solid state laser". It is the only ophthalmic laser that operates at 590nm for any eye surgery application.
- The CAPSULaser cataract system has been granted multiple US and worldwide patents for its unique design and functionality:
- CAPSULASER is a Selective Laser Capsulotomy (SLC) device that is absorbed by the proprietary CAPSULBlue dye that is placed in the capsule. The energy is 97% absorbed by the dye so that no energy gets to other critical eye anatomy.
- The capsule is stained with this patented trypan blue (CAPSULBlue 0.4% microfiltered) ophthalmic solution. This dye also been proven to provide a landmark that indicates the center of the visual axis with no other technology necessary. The surgeon can center the capsulotomy on the visual axis using this Trypan Capsular Landmark, this approach is also patented.
- While creating the capsulotomy it converts type-4 collagen to amorphous collagen which creates a stronger capsulotomy edge versus manual and Femto capsulotomies.
- In addition to the stronger capsulotomy edge, the CAPSULaser creates capsulotomies with more precise openings—circularity and diameter. This leads to 100% coverage of the edge of the IOL.
- Laser treatment time is 0.33 sec for a full circle at 5.0mm capsulotomy.
- Capsulotomy diameter range is in between 4.0mm to 5.5mm with 0.1 incremental steps.
- The Laser system is very compact and fits on most standard ophthalmic microscopes. It is currently compatible with Carl Zeiss, HAAG-STREIT, Topcon, & Leica microscopes with a 200 mm objective lens.
- This laser is very useful to create capsulotomies in complex cases like Intumescence and white cataract.

No other company/manufacturers have a device with the above features.
M/S Myhealthskape Medicals Pvt. Ltd. is exclusive distributor for CAPSULaser in India.

Thank you
Sincerely yours

David Mordaunt, CEO, EXCEL-LENS, Inc.



EXCEL-LENS, Inc., 23116 Summit Road, Los Gatos, CA 95033



12th June 2024

“LETTER OF AUTHORIZATION”

We, **EXCEL-LENS, Inc.**, having registered address at EXCEL-LENS, Inc., 2316 Orchard Parkway, Tracy, CA 95377, USA, **Contact no. +61 408 086 173** are reputed manufacturers of Ophthalmic medical devices.

We hereby certify that, **M/s. My Healthskape Medicals Private Limited, (“My Healthskape”)** having registered address at 4/425A, Sharma Industrial Estate, Walbhat Road, Goregaon East, Mumbai - 4000063, Maharashtra, India is our Exclusive Distributor in India for all the product range manufactured by us.

My Healthskape or any of this branch office is authorized to bid on equipment to be supplied to any government organization and other parties interested in equipment mentioned in the list of products.

No Company or Firm or Individual other than **My Healthskape** is authorized to bid, negotiate and conclude any contract in regards to business for equipment mentioned in the list of products in the entire territory of India.


My Healthskape is also authorized to provide after sales services for all the existing and new equipment installed.

This letter of authorization is valid till **31st December 2026** and shall be deemed to have been extended, until and unless terminated by either party.

Thanking you.

For **M/s. EXCEL-LENS, Inc.**,

Lee Walker-Roberts
Vice President of Sales
www.capsulaser.com
lwalkerroberts@capsulaser.com



CAPSULASER®
Selective Laser Capsulotomy

EXCEL-LENS, Inc., 455 North Canyons Parkway, Suite B, Livermore, CA 94551

List of EXCEL-LENS, Inc Devices

We, **EXCEL-LENS, Inc.**, having registered address at EXCEL-LENS, Inc., 2316 Orchard Parkway, Tracy, CA 95377, USA, **Contact no. +61 408 086 173** are reputed manufacturers of Ophthalmic medical devices hereby certify that, below list is product range which **M/s. My Healthskape Medicals Private Limited** having registered address at 4/425A, Sharma Industrial Estate, Walbhat Road, Goregaon East, Mumbai - 4000063, Maharashtra, India is allowed to distribute in India.

1. CAPSULaser
2. CAPSULBlue
3. CAPSULVisc
4. CAPSULGel
5. All other accessories associated with the proper function of CAPSULaser laser system.

This letter of authorization is valid unit 31st December 2026.



Lee Walker-Roberts
Vice President of Sales
www.capsulaser.com
lwalkerroberts@capsulaser.com



CAPSULASER®

Selective Laser Capsulotomy

Enclosed along – List of Products

EXCEL-LENS, Inc., 455 North Canyons Parkway, Suite B, Livermore, CA 94551

A draft of tender/Bid document submitted to see if formal approval is needed to publish on GeM portal.
Submitted please



A.S.O. (Surgical Spec, RPC)

Bid Number/क्रमांक (बिड संख्या):
GEM/2025/B/6108408
Dated/दिनांक: 03-04-2025

S.O. (Surgical Spec, RPC)
Bid Document/ बिड दस्तावेज

| Bid Details/बिड विवरण | |
|--|--|
| Bid End Date/Time/बिड बंद होने की तारीख/समय | 05-05-2025 12:00:00 |
| Bid Opening Date/Time/बिड खुलने की तारीख/समय | 05-05-2025 12:30:00 |
| Bid Offer Validity (From End Date)/बिड पेशकश वैधता (बंद होने की तारीख से) | 180 (Days) |
| Ministry/State Name/मंत्रालय/राज्य का नाम | Ministry Of Health And Family Welfare |
| Department Name/विभाग का नाम | Department Of Health And Family Welfare |
| Organisation Name/संगठन का नाम | All India Institute Of Medical Sciences (aiims) |
| Office Name/कार्यालय का नाम | Aiims, New Delhi |
| क्रेता ईमेल/Buyer Email | dharampal@aiims.gov.in |
| Total Quantity/कुल मात्रा | 2 |
| Item Category/मद केटेगरी | Selective Laser Capsulotomy System |
| GeMARPTS में खोजी गई स्ट्रिंग / Searched Strings used in GeMARPTS | Selective Laser Capsulotomy System |
| GeMARPTS में खोजा गया परिणाम / Searched Result generated in GeMARPTS | Selective Laser Trabeculoplasty (SLT) Lens, electronic total station, Holmium Laser, 3D Laser Scanning and imaging system (Forensic Use) |
| अधिसूचना के लिए चयनित प्रासंगिक श्रेणियाँ / Relevant Categories selected for notification | <ul style="list-style-type: none"> Digital Photo Slit Lamp with Applanation Tonometer Ophthalmic ND Yag laser Ophthalmic Operating Microscopes |
| MSE Exemption for Years Of Experience/अनुभव के वर्षों से एमएसई छूट/ and Turnover/टर्नओवर के लिए एमएसई को छूट प्राप्त है | Yes |
| Startup Exemption for Years Of Experience/अनुभव के वर्षों से स्टार्टअप छूट/ and Turnover/ टर्नओवर के लिए स्टार्टअप को छूट प्राप्त है | Yes |
| Document required from seller/विक्रेता से मांगे गए दस्तावेज | Certificate (Requested in ATC), OEM Authorization Certificate, Compliance of BoQ specification and supporting document *In case any bidder is seeking exemption from Experience / Turnover Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer |

Bid Details/बिड विवरण

| | |
|--|-----------------------------------|
| Do you want to show documents uploaded by bidders to all bidders participated in bid?/ | Yes |
| Bid to RA enabled/बिड से रिवर्स नीलामी सक्रिय किया | Yes |
| RA Qualification Rule | H1-Highest Priced Bid Elimination |
| Comprehensive Maintenance Charges Required | Yes |
| Type of Bid/बिड का प्रकार | Two Packet Bid* |
| Time allowed for Technical Clarifications during technical evaluation/तकनीकी मूल्यांकन के दौरान तकनीकी स्पष्टीकरण हेतु अनुमत समय | 5 Days |
| Inspection Required (By Empanelled Inspection Authority / Agencies pre-registered with GeM) | No |
| Evaluation Method/मूल्यांकन पद्धति | Total value wise evaluation |
| Financial Document Required/वित्तीय दस्तावेज की आवश्यकता है। | Yes |
| Arbitration Clause | No |
| Mediation Clause | No |

EMD Detail/ईएमडी विवरण

| | |
|-----------------------------|---------------------|
| Advisory Bank/एडवाइजरी बैंक | State Bank of India |
| EMD Amount/ईएमडी राशि | 660000 |

ePBG Detail/ईपीबीजी विवरण

| | |
|--|---------------------|
| Advisory Bank/एडवाइजरी बैंक | State Bank of India |
| ePBG Percentage(%) / ईपीबीजी प्रतिशत (%) | 5.00 |
| Duration of ePBG required (Months) / ईपीबीजी की अपेक्षित अवधि (महीने). | 26 |

(a). EMD EXEMPTION: The bidder seeking EMD exemption, must submit the valid supporting document for the relevant category as per GeM GTC with the bid. Under MSE category, only manufacturers for goods and Service Providers for Services are eligible for exemption from EMD. Traders are excluded from the purview of this Policy./जेम की शर्तों के अनुसार ईएमडी छूट के इच्छुक बिडर को संबंधित केटेगरी के लिए बिड के साथ वैध समर्थित दस्तावेज प्रस्तुत करने हैं। एमएसई केटेगरी के अंतर्गत केवल वस्तुओं के लिए विनिर्माता तथा सेवाओं के लिए सेवा प्रदाता ईएमडी से छूट के पात्र हैं। व्यापारियों को इस नीति के दायरे से बाहर रखा गया है।

(b). EMD & Performance security should be in favour of Beneficiary, wherever it is applicable./ईएमडी और संपादन जमानत राशि, जहां यह लागू होती है, लाभार्थी के पक्ष में होनी चाहिए।

Beneficiary/लामार्थी :

Chief, Dr. Rajendra Prasad Centre For Ophthalmic Sciences
Aiiims, New Delhi, Department of Health and Family Welfare, All India Institute of Medical Sciences (AIIMS),
Ministry of Health and Family Welfare
(Chief, Dr. Rajendra Prasad Centre For Ophthalmic Sciences)

MII Purchase Preference/एमआईआई खरीद वरीयता

| | |
|--|-----|
| MII Purchase Preference/एमआईआई खरीद वरीयता | Yes |
|--|-----|

MSE Purchase Preference/एमएसई खरीद वरीयता

| | |
|---|-----|
| MSE Purchase Preference/एमएसई खरीद वरीयता | Yes |
|---|-----|

1. If the bidder is a Micro or Small Enterprise as per latest definitions under MSME rules, the bidder shall be exempted from the requirement of "Bidder Turnover" criteria and "Experience Criteria" subject to meeting of quality and technical specifications. If the bidder is OEM of the offered products, it would be exempted from the "OEM Average Turnover" criteria also subject to meeting of quality and technical specifications. In case any bidder is seeking exemption from Turnover / Experience Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer.

2. If the bidder is a Startup, the bidder shall be exempted from the requirement of "Bidder Turnover" criteria and "Experience Criteria" subject to their meeting of quality and technical specifications. If the bidder is OEM of the offered products, it would be exempted from the "OEM Average Turnover" criteria also subject to meeting of quality and technical specifications. In case any bidder is seeking exemption from Turnover / Experience Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer.

3. Preference to Make In India products (For bids < 200 Crore): Preference shall be given to Class I local supplier as defined in public procurement (Preference to Make in India), Order 2017 as amended from time to time and its subsequent Orders/Notifications issued by concerned Nodal Ministry for specific Goods/Products. The minimum local content to qualify as a Class 1 local supplier is denoted in the bid document. If the bidder wants to avail the Purchase preference, the bidder must upload a certificate from the OEM regarding the percentage of the local content and the details of locations at which the local value addition is made along with their bid, failing which no purchase preference shall be granted. In case the bid value is more than Rs 10 Crore, the declaration relating to percentage of local content shall be certified by the statutory auditor or cost auditor, if the OEM is a company and by a practicing cost accountant or a chartered accountant for OEMs other than companies as per the Public Procurement (preference to Make-in -India) order 2017 dated 04.06.2020. Only Class-I and Class-II Local suppliers as per MII order dated 4.6.2020 will be eligible to bid. Non - Local suppliers as per MII order dated 04.06.2020 are not eligible to participate. However, eligible micro and small enterprises will be allowed to participate. The buyers are advised to refer the OM No.F.1/4/2021-PPD dated 18.05.2023.

OM No.1 4 2021 PPD dated 18.05.2023 for compliance of Concurrent application of Public Procurement Policy for Micro and Small Enterprises Order, 2012 and Public Procurement (Preference to Make in India) Order, 2017.

4. Purchase preference will be given to MSEs having valid Udyam Registration and whose credentials are validated online through Udyam Registration portal as defined in Public Procurement Policy for Micro and Small Enterprises (MSEs) Order, 2012 dated 23.03.2012 issued by Ministry of Micro, Small and Medium Enterprises and its subsequent Orders/Notifications issued by concerned Ministry. If the bidder wants to avail themselves of the Purchase preference, the bidder must be the manufacturer / OEM of the offered product on GeM. Traders are excluded from the purview of Public Procurement Policy for Micro and Small Enterprises and hence resellers offering products manufactured by some other OEM are not eligible for any purchase preference. In respect of bid for Services, the bidder must be the Service provider of the offered Service. Relevant documentary evidence in this regard shall be uploaded along with the bid in respect of the offered product or service and Buyer will decide eligibility for purchase preference based on documentary evidence submitted, while evaluating the bid. If L-1 is not an MSE and MSE Seller (s) has / have quoted price within L-1+ 15% (Selected by Buyer) of margin of purchase preference /price band defined in relevant policy, such MSE Seller shall be given opportunity to match L-1 price and contract will be awarded for 25% (selected by Buyer) percentage of total quantity. The buyers are advised to refer the OM No. F.1/4/2021-PPD dated 18.05.2023 OM No.1 4 2021 PPD dated 18.05.2023 for compliance of Concurrent application of Public Procurement Policy for Micro and Small Enterprises Order, 2012

and Public Procurement (Preference to Make in India) Order, 2017. Benefits of MSE will be allowed only if seller is validated on-line in GeM profile as well as validated and approved by Buyer after evaluation of documents submitted.

5. Reverse Auction would be conducted amongst all the technically qualified bidders except the Highest quoting bidder. The technically qualified Highest Quoting bidder will not be allowed to participate in RA. However, H-1 will also be allowed to participate in RA in following cases:

- If number of technically qualified bidders are only 2 or 3.
- If Buyer has chosen to split the bid amongst N sellers, and H1 bid is coming within N.
- In case Primary product of only one OEM is left in contention for participation in RA on elimination of H-1.
- If L-1 is non-MSE and H-1 is eligible MSE and H-1 price is coming within price band of 15% of Non-MSE L-1.
- If L-1 is non-MII and H-1 is eligible MII and H-1 price is coming within price band of 20% of Non-MII L-1.

Pre Bid Detail(s)

| Pre-Bid Date and Time | Pre-Bid Venue |
|-----------------------|--|
| 15-04-2025 16:00:00 | Room No. 180, First Floor, Store Section, AIIMS, New Delhi |

Selective Laser Capsulotomy System (2 pieces)

(Minimum 50% and 25% Local Content required for qualifying as Class 1 and Class 2 Local Supplier respectively/क्रमशः श्रेणी 1 और श्रेणी 2 के स्थानीय आपूर्तिकर्ता के रूप में अर्हता प्राप्त करने के लिए आवश्यक)

Technical Specifications/तकनीकी विशिष्टियाँ

| | |
|---|--------------------------|
| Buyer Specification Document/क्रेता विशिष्टि दस्तावेज | Download |
|---|--------------------------|

| Comprehensive Maintenance | |
|--|--------|
| Warranty of required product | 2 Year |
| Comprehensive Maintenance Duration (Post Warranty) | 8 Year |

*Warranty displayed under the AMC/CMC Details section will supersede the warranty displayed under the catalog specification

Consignees/Reporting Officer/परेषिती/रिपोर्टिंग अधिकारी and/ तथा Quantity/मात्रा

| S.No./क्र. सं. | Consignee Reporting/Officer/परेषिती/रिपोर्टिंग अधिकारी | Address/पता | Quantity/मात्रा | Delivery Days/डिलीवरी के दिन |
|----------------|--|--------------------------------------|-----------------|------------------------------|
| 1 | Dharampal Singh | 110029,AIIMS, Ansari Nagar New Delhi | 2 | 90 |

Buyer Added Bid Specific Terms and Conditions/क्रेता द्वारा जोड़ी गई बिड की विशेष शर्तें

1. Experience Certificate for the supply of the same to any Govt/ PSU/ any renowned private organisation along with Supply/ Purchase Order.
2. If the agency is registered under MSME or NSIC, then EMD exemption certificate needs to be enclosed.
3. Make in india specific authorisation certificate needs to be enclosed.
4. **Buyer Added Bid Specific ATC**

Buyer uploaded ATC document [Click here to view the file.](#)

Disclaimer/अस्वीकरण

The additional terms and conditions have been incorporated by the Buyer after approval of the Competent Authority in Buyer Organization, whereby Buyer organization is solely responsible for the impact of these clauses on the bidding process, its outcome, and consequences thereof including any eccentricity / restriction arising in the bidding process due to these ATCs and due to modification of technical specifications and / or terms and conditions governing the bid. If any clause(s) is / are incorporated by the Buyer regarding following, the bid and resultant contracts shall be treated as null and void and such bids may be cancelled by GeM at any stage of bidding process without any notice:-

1. Definition of Class I and Class II suppliers in the bid not in line with the extant Order / Office Memorandum issued by DPIIT in this regard.
2. Seeking EMD submission from bidder(s), including via Additional Terms & Conditions, in contravention to exemption provided to such sellers under GeM GTC.
3. Publishing Custom / BOQ bids for items for which regular GeM categories are available without any Category item bunched with it.
4. Creating BoQ bid for single item.
5. Mentioning specific Brand or Make or Model or Manufacturer or Dealer name.
6. Mandating submission of documents in physical form as a pre-requisite to qualify bidders.
7. Floating / creation of work contracts as Custom Bids in Services.
8. Seeking sample with bid or approval of samples during bid evaluation process. (However, in bids for attached categories, trials are allowed as per approved procurement policy of the buyer nodal Ministries)
9. Mandating foreign / international certifications even in case of existence of Indian Standards without specifying equivalent Indian Certification / standards.
10. Seeking experience from specific organization / department / institute only or from foreign / export experience.
11. Creating bid for items from irrelevant categories.
12. Incorporating any clause against the MSME policy and Preference to Make in India Policy.
13. Reference of conditions published on any external site or reference to external documents/clauses.
14. Asking for any Tender fee./ Bid Participation fee / Auction fee in case of Bids / Forward Auction, as the case may be.

Further, if any seller has any objection/grievance against these additional clauses or otherwise on any aspect of this bid, they can raise their representation against the same by using the Representation window provided in the bid details field in Seller dashboard after logging in as a seller within 4 days of bid publication on GeM. Buyer is duty bound to reply to all such representations and would not be allowed to open bids if he fails to reply to such representations:

Additional Clause For Comprehensive Maintenance Charges

1.CMC shall include preventive maintenance including calibration as per technical/ service /operational manual of the manufacturer, service charges and spares, after satisfactory completion of Warranty. During the CMC period commencing from date of the successful completion of warranty period, Service personnel shall visit each consignee site as recommended in the manufacturer's technical/ service /operational manual, at least once in six

months or as per user requirement. Cost of consumables shall not be included in CMC. Further there will be 98% uptime warranty during CMC period on 24 (hrs) X 7 (days) X 365 (days) basis, with penalty, to extend CMC period by double the downtime period.

2. CMC charges to be indicated as percentage of cost of equipment quoted for each year after the warranty period.

3. GST shall be included in the CMC Charges quoted.

4. Cost of CMC will be added for Ranking/Evaluation purpose with depreciation formula. A 10% discounting rate per year shall be applied on CMC Charges for price evaluation on Net Present Value.

5. The payment of CMC will be made on quarterly basis after satisfactory completion of said period, duly certified by end user.

6. While creating a bid or RA, buyers shall indicate whether CMC is required against Yes/No options. If CMC Charges are included, an option for number of years for CMC required after the warranty period shall be available. Under this option up to 10 years can be chosen for CMC charges beyond warranty period.

7. In case the bid has a provision for CMC, the warranty of the product will also be deemed to have been converted into Comprehensive warranty including preventive maintenance and calibration as per technical/ service /operational manual of the manufacturer, service charges and spares, during the Warranty Period also. Sellers are therefore advised to include the cost of Comprehensive Warranty including spares (excluding consumables) also in product Cost.

8. The CMC functionality shall be available in bid only and no direct RA shall be applicable. In case of bid to R/A decrement rules shall be applicable on total price inclusive of CMC charges. Bunching of products shall not be available while creating bids with CMC charges.

8.1. Buyer shall indicate number of years of warranty by selecting different options available in the field depending on warranty parameter applicable in category parameters for the equipment. No. of years of warranty indicated here shall supersede the warranty period indicated elsewhere in bid or product specifications. The Seller while participating in Bid/RA will get fields to indicate CMC charges as percentage depending on number of years of CMC selected by Buyer. The following shall be applicable, if 5 year CMC selected:

CMC charges for 1st year after warranty period- Percentage to be indicated- A1

CMC charges for 2nd year after warranty period- Percentage to be indicated- A2

CMC charges for 3rd year after warranty period - Percentage to be indicated- A3

CMC charges for 4th year after warranty period - Percentage to be indicated- A4

CMC charges for 5th year after warranty period - Percentage to be indicated- A5

Similarly, A6 to A10 are to be indicated for 6th to 10th year of CMC if applicable.

8.2. The calculation of CMC Charges shall take into account the number of years of warranty and duration of CMC as specified while creating bid.

8.3. In the price evaluation, the system shall provide function to calculate the cost of each equipment by formula indicated below including CMC and then show the inter-se-ranking of the bidders. The following are the variables

(i) Number of years for which CMC required.

(ii) Number of years of product warranty

The formula for calculating total cost including CMC charges shall be as under:

Total Cost for evaluation =

$C + C * \{ (A1/100)/(1.10^n) + (A2/100)/(1.10^{n+1}) + (A3/100)/(1.10^{n+2}) + (A4/100)/(1.10^{n+3}) + (A5/100)/(1.10^{n+4}) \}$ and so on

C - Cost for equipment quoted and n shall be number of years of product warranty specified.

If 2 year warranty specified, n shall be 2 and if 5 year warranty specified, n shall be 5. A1, A2, A3, A4 & A5 shall depend on how many years CMC selected. For 3 year CMC, only A1, A2 and A3 factors are to be taken into account and A4 and A5 will not be applicable.

8.4. CMC charges offered for each subsequent year should be same or higher than preceding year.

8.5. The CMC charges shall be offered within range of 3 to 10% of cost of equipment.

9. Since CMC charges are to be paid only later for each year during CMC period, applicable performance guarantee amount after placement of contract shall be based on the cost of equipment excluding the cost of CMC Charges.

10. Performance bank guarantee applicable for CMC is to be submitted at start of the CMC and shall be applicable between 2.5% to 10% as specified in bid on total CMC Charges. The PBG submitted after award of contract shall be released only after new PBG for the CMC period is submitted and accepted by buyer/consignee after due verification. Bank guarantee for CMC is to remain valid till completion of CMC period plus one year. The bank guarantee for CMC shall be submitted to buyer directly. In case, seller fails to submit the PBG or does not provide services for the CMC contract after expiry of warranty period then PBG of equipment shall be forfeited.

11. In case of splitting of order quantity, equipment cost and CMC charges offered by L1 bidder shall be matched by higher quoting eligible bidders on one-to-one basis. The equipment cost and CMC charges (year to year) shall be matched individually.

12. The CMC Contract shall be an offline contract to be handled by buyer. The payment of CMC will be made on quarterly basis after satisfactory completion of said period, duly certified by end user and scope of CMC will be as per para 1 above.

13. CMC Charges are inclusive of all the charges for Transportation, Lodging, Boarding, all insurances including third party insurance and all other incidental charges. The same shall include GST. The prices also include cost of spares and damaged parts. Purchaser does not have any liability, whatsoever, over and above the cost of CMC. It also includes for arranging hand tools & tackles, special tools etc. required to carry out the work.

All GeM Sellers / Service Providers are mandated to ensure compliance with all the applicable laws / acts / rules including but not limited to all Labour Laws such as The Minimum Wages Act, 1948, The Payment of Wages Act, 1936, The Payment of Bonus Act, 1965, The Equal Remuneration Act, 1976, The Payment of Gratuity Act, 1972 etc. Any non-compliance will be treated as breach of contract and Buyer may take suitable actions as per GeM Contract.

This Bid is also governed by the General Terms and Conditions/ यह बिड सामान्य शर्तों के अंतर्गत भी शासित है

In terms of GeM GTC clause 26 regarding Restrictions on procurement from a bidder of a country which shares a land border with India, any bidder from a country which shares a land border with India will be eligible to bid in this tender only if the bidder is registered with the Competent Authority. While participating in bid, Bidder has to undertake compliance of this and any false declaration and non-compliance of this would be a ground for immediate termination of the contract and further legal action

In accordance with the laws, जेम की सामान्य शर्तों के खंड 26 के संदर्भ में भारत के साथ भूमि सीमा साझा करने वाले देश के बिडर से खरीद पर प्रतिबंध के संबंध में भारत के साथ भूमि सीमा साझा करने वाले देश का कोई भी बिडर इस नियम में बिड देने के लिए तभी पात्र होगा जब वह बिड देने वाला सक्षम प्राधिकारी के पास पंजीकृत हो। बिड में भाग लेते समय बिडर को इसका अनुपालन करना होगा और कोई भी गलत घोषणा किए जाने व इसका अनुपालन न करने पर अनुबंध को तत्काल समाप्त करने और कानून के अनुसार आगे की कानूनी कार्रवाई का आधार होगा।

---Thank You/धन्यवाद---

Technical Specifications of “Selective Laser Capsulotomy System – 02 Nos.”

1. The Laser type should be “Optically pumped thin disc solid laser”.
2. The wavelength of the Laser should be 590 ± 3 nm.
3. The colour of Laser should be Orange colour.
4. The nominal output power should be ≤ 3500 nW.
5. Duty cycle should be 100% continuous waves.
6. Divergence should be 0.088 NA Full angle.
7. The Laser should be use exclusively to make Capsulotomy only.
8. Laser treatment time should be 0.33 sec for a full circle for 5.00 mm Capsulotomy.
9. Capsulotomy diameter range should be between 4.0 mm to 5.5 mm with 0.1 incremental step.
10. Laser should create rolled over edge elastic Capsulotomy.
11. Laser console should have Reticle intensity adjustment knob.
12. Laser should come with eye safety filter.
13. Laser should come with dedicated footswitch.
14. Laser should come with patient fixation ring to hold the eye during treatment.
15. Laser console design should be compact.
16. Laser delivery system should be mounted on microscopes & should be compatible with Carl Zeiss, Hagg Streit & Leica microscopes with 200 mm Objective Lens only.
17. The maximum weight of laser delivery system should be less than 2 kg.
18. Laser console and laser delivery system should be connected through high quality and sturdy optical fibre cable.
19. Classification per IEC 60825-1: 2014: Class 4
20. Aiming beam type should be Direct Laser Diode.
21. Wavelength of Aiming beam should be 630 to 645 nm.
22. Aiming beam Power output : From a minimum that is barely visible to a maximum that is less than < 10 mW.
23. Aiming beam laser Classification per IEC 60825-1: 2014 Class 2.
24. Electric requirement : Voltage 120-230 V 3.0A 50/60Hz & Fusing 250V T 3.0A.
25. Laser operating temperature $15^{\circ} - 32^{\circ} \text{C}$ ($59^{\circ} - 90^{\circ} \text{F}$)
26. Maximum humidity : Up to 90% at 32°C (90°F) non condensing.
27. Following consumables should be supply along with each Laser system as follows:-
 - CapsulBlue with cannula (Microfiltered Trypan blue (0.4%) ophthalmic solution – 1000 Nos.
 - Patient guide/ fixation ring – 100 Nos.
 - Procedure license key (100 Nos.) for 1000 Procedures.
 - Laser should come with proprietary CapsulBlue (Trypan Blue) to stain the anterior capsule.
28. Laser should be CE/FDA/CDSO