

Department of Reproductive Biology

✓ Prof. Incharge
Computer Facility
AIIMS

Subject: Request to download for inviting quotations.

It is requested that the quotations of the pipette and bio-safety cabinet as given below pertaining the following matter may kindly be download on AIIMS concerned website as early as possible.

The quotations of the above said "**Digital Autoclavable Pipette and one small bio-safety cabinet for molecular assembling work**" should be reach the office of the Department of Reproductive Biology, room No. 2083, IInd floor, Teaching Block, AIIMS, New Delhi latest by 20th Feb. 2014. Please send the manufacturer/principal's authority letter as well as the names & contact numbers of AIIMS/Govt. departments supplied by the supplier/dealer's, in a sealed envelope. All quotations should be typed written or ink.

The request has already been mailed to you by our store keeper.



Dr. Anand Kumar MD, FAMS
Prof. & Head
Head of the Department of Reproductive Biology
All India Institute of Medical Sciences
New Delhi-110029, India

Department of Reproductive Biology

Rate-Enquiry No. 05/R.B/CRIA/2013-14

You are requested to send the separate quotations for- "**Digital Autoclavable Pipette and one small bio-safety cabinet for molecular assembling work**" through Courier/post.

- 1. Small bio-safety cabinet** **Quantity 1**
- Size (W/D/H):** 24" x 24" x 30" (± 2 ")
- UV sterilization:** Built-in ~254 nm ultraviolet light for decontamination of consumables and equipment between amplifications
UV lamp only can be switched on when the front window is closed
UV timer
UV light (4-5 nos.)
- Laminar flow:** Vertical laminar flow action to prevent outside contaminants from entering the work area during manipulation
HEPA filter with efficiency of over 99% at 0.3um
Airflow speed: 0.3-0.5m/s
Pre-filter (10 nos.)
- Main Body:** Steel plate with anti-bacteria coating
- Work Table:** Custom extruded polypropylene blend that reflects ultraviolet energy and makes cleaning easier
- Shelf:** Removable shelf with pipettor holder
- Sockets:** Power cord access port (two)
- Caster:** Universal wheel

General Terms & Conditions

The Vendor should have a good service and application support back up with user list
Warranty for 2 Years or more

- 2. Pipette set: Digital & Autoclavable** **Quantity 1 set (4 pcs)**
- Volume:** ≤ 2 ul, ≤ 20 ul, ≤ 200 ul & ≤ 1000 ul (one each) pipette
- Battery & Electric power operated
- Charging strand:** one for each pipette
- Fully Autoclavable and UV resistant
- Digital display
- Piston:** smooth running, corrosion-free and chemically resistant
- Easy to maintain & calibratable
- Should use tips from all well known manufacturers

General Terms & Conditions

The Vendor should have a good service and application support back up with user list

Warranty for 2 Years or more



Prof. Anand Kumar MD, FAMS
Head of the Department of Reproductive Biology
All India Institute of Medical Sciences
New Delhi-110029, India