

DEPARTMENT OF HEMATOLOGY
A.I.I.M.S. NEW DELHI

18.11.2015

CORRIGENDUM

To

Dr. P.Kotwal
Chairman
Computer Facility, AI.I.I.MS., New Delhi


Sub: For giving in website Rate Enquiry No. 3/Haem/2015-16 for purchase of Thermostat Hotplate

Dear Sir,

Enclosed the specification for Thermostat Hotplate . Invite the Sealed quotations for Thermostat Hotplate as per specifications given below. Quotations to be send to Dr. Renu Saxena, Prof. & Head, Department of Hematology, A.I.I.M.S. New Delhi 110029. The last date for receiving the sealed quotation is 4th December, 2015. Kindly include in AIIMS, Website.

Thanking you,

Yours Sincerely


Dr. RENU SAXENA
PROF. & HEAD, DEPT. HEMATOLOGY

DEPARTMENT OF HEMATOLOGY, A.I.I.M.S.

SPECIFICATION FOR THERMOSTAT HOT PLATE

Quantity : 1 no.

- Should have a ambient +5 UP to 200°C Stability +0.5°C
- Should have a Electronic digital temperature and time control display
- Should have suitable for thermal tests (not suitable for boiling tests)
- Should have surface plate coated with P.T.F.E. for tprotection against chemicals
- Should have a indicator alarm if the plate is over temperature
- Should have ISO Certificate.

It should be two bid system Technical bid & Price bid Should be put in separate sealed envelope.