

**DEPARTMENT OF RADIODIAGNOSIS  
ALL INDIA INSTITUTE OF MEDICAL SCIENCES  
NEW DELHI-29**

Short Rate Enquiry No. 7/2017-18/Radio-diagnosis/AIIMS

Dated: 23.2.2018

Purchase of Light Weight Lead Free Aprons : Qty: 15 Aprons

The sealed quotations are invited by post/ per bearer for the supply of Light Weight Lead Free Aprons for the Department of Radio-diagnosis, AIIMS, New Delhi-29.


The quotation duly signed should reach the office of Professor & Head, Deptt. of Radio-diagnosis, AIIMS, New Delhi-29 on or before 7.3.2018 at 5 PM. The rate should be valid for one year.

1. The Aprons must be manufactured from a reputed manufacturer and should be USA-FDA enlisted.
2. Product should be European CE marked and certified as meeting the requirements of EU directive 89/686/EEC from a notified body.
3. Offered product should satisfy the requirement of 'Protective devices against diagnostic medical X Radiation' based on EN 61331-3:2014. It should be certified from a notified body.
4. The copy of Quality certificate must be presented at the time of demonstration. The quoted model name and product code must be mentioned on all the presented certificates.
5. The Aprons must consist of lead free protective materials and manufactured with non toxic heavy metals.
6. Aprons must have lead equivalence of 0.5mm (Pb).  
The same must be mentioned on the apron.
7. The size of the apron must be 60-65cm in width and 95-100cm in Length
8. The Aprons must have the attenuation report certifications from AERB.
9. The supplier must provide AIIMS Logo and department name on the aprons.
10. The apron must be ultra light.
11. The original manufacturer must give Guarantee for 02 (Two) years for any manufacture defect in core material.
12. The Aprons should have an integrated (attached) thyroid collar for additional protection of thyroid area so that no separate thyroid shield is required.
13. The Aprons should have stuffed shoulder pads (anti skid pads) to offer the maximum comfort over the shoulders.
14. The Aprons must have special sweat absorbing fabric which should prevent perspiration to increase comfort even in warm conditions.
15. Stains of blood, contrast etc should be easily washable.
16. Demonstration : a departmental committee will examine and assess the aprons submitted by the vendors. Decision of the committee will be final. Date and time of demonstration will be communicated to the participating vendors.

The firm should submit the following:

- Price reasonable certificate
- An undertaking on the official letter head of the company that firm has not been black listed.

**Note:** The quotation of technically selected samples will be open for price bid.

  
Dr. Arun Kumar Gupta, MD, FAMS  
Professor & Head  
Department of Radiodiagnosis  
All India Institute of Medical Sciences  
New Delhi-110029