

Magnetic Stirrer with heating options

Specifications:

1. Small size
2. Turn knobs for temperature and speed control
3. Option for heating
4. Safety indicator LED
5. Power 5A/600W
6. Speed 60 to 1,200rpm
7. Temperature Range (Metric) 38° to 220°C
8. Surface Area (Metric) Heating not less than 22 cm²
9. Voltage 120V Amperage 8.9A Hertz 50/60Hz Wattage 1070w
10. Height ≤20.8cm Length≤32.4cm Width ≤25.4cm
11. Certifications/Compliance ISO 9008
12. Description 120V 50/60Hz, 1070w, 8.9A
13. Magnetic Stirrer beads to be provided- length- 1cm, 2.5cm, 5cm, 9cm. Width: 0.75-1.5 cm
14. Copies of all certifications e.g. Quality Standard certificate, Propriety Item/parts, Patent of parts/ technology, Principal company/Authorised Distributorship should be attached with the quotations.
15. In-House Service Engineers from principal company/ or authorised agents should be available on one-day notice basis in case of emergency.
16. Only Principal Companies or Authorized Distributors from Principal companies should quote. Quotations from non - authorized distributors will not be entertained.
17. Compliance sheet should be attached along with the quotation. Compliant points should be highlighted in the company catalogue. Compliant points should be given (in the sheet) in order of the specifications' serial order.

Devdancer Ray

Ashraf

डॉ० ए. शरीफ/Dr. A. SHARIFF
आचार्य/Professor
शरीर रचना विभाग/Deptt. of Anatomy
अ.भा.आ.सं., नई दिल्ली/AIIMS, New Delhi-29

Renu Dhingra
Dr. Renu Dhingra
Professor
Department of Anatomy
New Delhi-29



डॉ० टोनी जॉर्ज जैकब
Dr. TONY GEORGE JACOB MD, Ph.D.
सहायक आचार्य/Assistant Professor
शरीर रचना विभाग/Deptt. of Anatomy
अ.भा.आ.सं., नई दिल्ली-29/A.I.I.M.S., New Delhi-29

[Signature]

Terms & Conditions:

- The quotations should be submitted to the Professor & Head, Department of Anatomy, Room No. 1010, AIIMS, New Delhi-29 and sent by post/per bearer by 7th October, 2016 up to 12.00 P.M. Quotation should be sealed in an envelope and the "name of the item" clearly written on top of the envelop.
- The make of the article offered should invariably be quoted. Quotation should be typed written in ink. All overwriting and erased entries will be deleted from the quotation. The rates should be valid for at least THREE MONTHS.
- VAT/Sales tax or any other kind of tax(s) must be separately mentioned against each item. In case no sales tax is chargeable, prices must be quoted as NET.
- **Warranty/Guarantee:** two years followed by three years CMC comprehensive warranty.
- **Installation:** The firm install the equipment on the site, free of cost
- The firm should be able to give demonstration of the item if required within 5 days of requisition.
- The firm must supply the item within 20 days after confirmation of order.
- **The successful bidder has to submit performances bank guarantee of 10% of the order value valid upto warranty period.**
- The payment will be made electronically viz RTGS/NEFT against delivery and satisfactory installation of the item. The following information should be also mentioned in the quotation.
- Name of the beneficiary:
- Account No. of the beneficiary:
- IFCS code of the bank/branch:
- The quantity can be increased or decreased to any extent depending upon the actual requirement.
- The Professor & Head of the Department reserves the right to cancel/reject full or any part of the rate enquiry which generally do not fulfil the conditions stipulated in the rate enquiry.
- Payment will be made after receiving the items in good condition and satisfactory installation.
- No extra transportation charge will be given to the firm.
- The firm must supply the items within the date mentioned on the supply order copy.

Thanking you

Renu Dhingra
(Dr. Renu Dhingra)

Professor

Computer facility: With request for uploading the same in AIIMS websites for the period of 15 Days.

Dr. Renu Dhingra
Professor
Department of Anatomy
A.I.I.M.S., New Delhi-29